



Intraoperative continuous noninvasive hemoglobin monitoring in patients with placenta previa undergoing cesarean section: a prospective observational study

Hyunsung Kim, Sang-Hwan Do, Jung-Won Hwang, and Hyo-Seok Na

Department of Anesthesiology and Pain Medicine, Seoul National University Bundang Hospital, Seongnam, Korea

Received February 15, 2019

Revised April 9, 2019

Accepted April 25, 2019

Corresponding author

Hyo-Seok Na, M.D., Ph.D.
Department of Anesthesiology
and Pain Medicine, Seoul National
University Bundang Hospital, 82
Gumi-ro 173beon-gil, Bundang-gu,
Seongnam 13620, Korea
Tel: 82-31-787-7507
Fax: 82-31-787-4063
E-mail: hsknana@gmail.com

ORCID

<https://orcid.org/0000-0003-0986-3243>

This paper was presented in the 22th Annual Meeting of Koreans Society of Obstetric Anesthesiologists.

Background: Obstetric patients with placenta previa are at risk for severe peripartum hemorrhage. Early detection of anemia and proper transfusion strategy are important for the management of obstetric hemorrhage. In this study, we assessed the utility and accuracy of noninvasive hemoglobin (SpHb) monitoring in patients with placenta previa during cesarean section.

Methods: Parturients diagnosed with placenta previa and scheduled for cesarean section under spinal anesthesia were enrolled. SpHb and laboratory Hb (Lab-Hb) were measured during surgery as primary outcomes.

Results: Seventy-four pairs of SpHb and Lab-Hb were collected from 39 patients. The correlation coefficient was 0.877 between SpHb and Lab-Hb ($P < 0.001$). The Bland-Altman plot showed a mean difference \pm SD of 0.3 ± 0.8 g/dl between noninvasive Hb and Lab-Hb, and the limits of agreement were -1.2 to 1.8 g/dl. The magnitude of the difference between SpHb and Lab-Hb was < 0.5 g/dl in 64.9%; however, it was > 1.5 g/dl in 10.8%.

Conclusions: SpHb monitoring had a good correlation with Lab-Hb. A small mean difference between SpHb and lab-Hb might not be clinically significant; however, the limits of agreements were not narrow. In particular, SpHb could be overestimated in the anemic population. Based on our results, further studies investigating the accuracy and precision of SpHb monitoring should be performed in parturients presenting Hb below 10 g/dl.

Keywords: Cesarean section; Laboratory hemoglobin; Noninvasive hemoglobin.

INTRODUCTION

Pregnant women present physiologic anemia during pregnancy, thus peripartum bleeding can cause more severe anemia during vaginal delivery or the cesarean section (C-sec). When parturients with placenta previa undergo C-sec, there may be increased chance of severe intraoperative bleeding, followed by profound anemia requiring massive blood trans-

fusion. For the purpose of frequent laboratory evaluation, proper transfusion strategy, and vital sign monitoring, intraoperative invasive arterial monitoring is usually performed.

Noninvasive hemoglobin (Hb) monitoring is useful when requiring continuous monitoring of Hb and performing proper transfusion; it is applicable in various clinical situations, such as managing critically ill patients with bleeding tendency, trauma patients with continuous bleeding, as well

as small pediatric patients with difficulty in frequent blood sampling, orthopedic operation with intraoperative substantial blood loss, or hemorrhagic shock [1–5]. However, this view remains contentious; some believe that noninvasive Hb monitoring may not be as accurate as the laboratory Hb, and thus asserting that it should not be used exclusively to guide transfusion [2,6,7].

During the peripartum period, it was found that noninvasive Hb monitoring was well correlated with laboratory Hb [8,9]. However, to the best of our knowledge, there is no study evaluating the usefulness of noninvasive Hb monitoring in parturient with placenta previa. When surgical bleeding occurs in parturients with placenta previa during C-sec, it is generally acute and extensively massive; therefore, transfusion is performed based on clinical implication until laboratory Hb results are obtained in such cases. If intraoperative Hb is monitored promptly and accurately by continuous noninvasive Hb monitoring in parturients with placenta previa, anemia can be detected instantly, which will allow proper fluid and transfusion management.

In this study, we applied continuous noninvasive Hb monitoring in parturients with placenta previa undergoing C-sec and compared the Hb level between continuous noninvasive Hb monitor and laboratory Hb level.

MATERIALS AND METHODS

This study was approved by the Institutional Review Board of Seoul National University Bundang Hospital (no. B-1706/403-306) and registered at ClinicalTrials.com (NCT 03409822).

Parturient women who underwent elective C-sec due to placenta previa were enrolled in this study. The exclusion criteria were as follows: American Society of Anesthesiology physical status 3 or more; refusal to participate in this study; peripheral vascular disease; cyanotic condition due to the congenital or acquired heart disease; refusal of transfusion; and inability to communicate in Korean. All participants provided written informed consent.

On arrival to the operating room, routine monitoring—such as electrocardiography, noninvasive blood pressure, and pulse oximetry—was initiated. In addition, invasive arterial pressure monitoring was also performed, and laboratory Hb (Lab-Hb) was evaluated using arterial blood via Stat

Profile pHox Ultra (Nova Biomedical, USA). Noninvasive Hb (SpHb) was measured by using the Radical-7 Pulse CO-oximeter (Masimo Corp., USA). Spinal anesthesia was performed using 7–9 mg of hyperbaric bupivacaine with 15 µg of fentanyl at lateral decubitus position, during which, 500 ml of colloid was routinely infused to prevent hypotension. Intraoperative hypotension was managed with phenylephrine or ephedrine, and bradycardia was treated with atropine. Spinal anesthesia was converted to general anesthesia if the operation was prolonged unexpectedly. Red blood cells (RBC) transfusion was initiated when Lab-Hb was below 8 g/dl, and the unit of RBC transfusion was decided according to the severity of surgical bleeding. SpHb was calibrated based on the initial lab-Hb before starting the operation. When the likelihood of massive and continued intraoperative bleeding was relatively high, evaluation of Lab-Hb and SpHb was at the discretion of the attending anesthesiologist.

The primary outcomes were SpHb and Lab-Hb. The secondary outcomes were perfusion index (Pi) measured simultaneously with SpHb by Radical-7 Pulse CO-oximeter, infused fluid volume, unit of transfusion, calculated blood loss during surgery, and urine output.

Statistical analysis

In previous studies [1,2,10,11], the range of correlation coefficient between SpHb and Lab-Hb was reported from 0.47 to 0.72. It was assumed that the correlation coefficient would be approximately 0.7 in our study. Adjusting for a 10% dropout, 39 patients were required to give a statistical power of 80% and a type 1 error of 5%.

The characteristics of patients and clinical factors were presented as the mean and standard deviation (SD) or number (%). Correlation between SpHb and Lab-Hb was examined by Spearman's rank correlation analysis. The differences between the two parameters were analyzed by one sample *t*-test. Bland–Altman analysis was used to assess the agreement between the two methods measuring Hb, with the limits of agreement defined as 2 SD from the mean. All statistical analyses were performed using IBM SPSS Statistics version 25.0 (IBM Corp., USA). P value of less than 0.05 was considered statistically significant.

RESULTS

The characteristics of 39 patients, surgery, and anesthesia were shown in Table 1. The mean amount of blood transfusion and incidence of patients receiving transfusion during surgery were presented in Table 2. The change of SpHb and Pi were presented at each time point during anesthesia (Fig.

Table 1. The Characteristics of Patients, Surgery, and Anesthesia

Variable	Value
Age (yr)	35.4 ± 3.0
Height (cm)	163.4 ± 4.5
Weight (kg)	68.5 ± 7.5
Operation time (min)	72.6 ± 22.9
Anesthesia time (min)	94.1 ± 23.4
Intraoperative crystalloid (ml)	802 ± 745
Intraoperative colloid (ml)	682 ± 291
Urine output (ml)	195 ± 223
Calculated intraoperative blood loss (ml)	967 ± 719

Values are presented as mean ± SD.

Table 2. Intraoperative Transfusion

Variable	Units	Incidence (n = 39)
RBC	0.9 ± 1.9	12 (30.8)
FFP	0.2 ± 0.9	3 (7.7)
Platelets	0.2 ± 1.0	1 (2.6)

Values are presented as mean ± SD or number (%). RBC: red blood cells, FFP: fresh frozen plasma.




Fig. 1. Change of noninvasive hemoglobin (SpHb) and perfusion index (Pi). OP: operation.

1). After colloid loading during spinal anesthesia, the mean SpHb decreased by 0.3 g/dl.

A total of 74 pairs of SpHb and Lab-Hb were collected from 39 patients. The correlation coefficient between SpHb and Lab-Hb was 0.877 ($P < 0.001$; Fig. 2). The Bland-Altman plot showed a mean difference ± SD of 0.3 ± 0.8 g/dl between non-invasive Hb and Lab-Hb, and the limits of agreement ranged from -1.2 to 1.8 g/dl (Fig. 3).

The magnitude of difference between SpHb and Lab-Hb was < 0.5 g/dl in 64.9%; however, its difference was 1.5 g/dl in 10.8% (Table 3). When the subgroups were classified according to Pi values, there were significantly more patients in the $Pi \geq 2.0$ subgroup presenting low discrepancy between SpHb




Fig. 2. Correlation between the laboratory hemoglobin (Lab-Hb) and noninvasive hemoglobin (SpHb). The correlation coefficient was 0.877 between the SpHb and Lab-Hb.




Fig. 3. Bland-Altman analysis for laboratory hemoglobin (Lab-Hb) and noninvasive hemoglobin (SpHb). The mean difference (one horizontal solid line) between the SpHb and Lab-Hb was 0.3 g/dl and the limits of agreement (two horizontal dash lines) were -1.2 to 1.8 g/dl.

Table 3. Absolute Value of the Difference between the Noninvasive and Laboratory Hemoglobin

SpHb-lab-Hb	< 0.5 g/dl	0.5–1.5 g/dl	> 1.5 g/dl
Total (n = 74)	48 (64.9)	18 (24.3)	8 (10.8)
Pi ≥ 2.0 (n = 48)	33 (44.6)	13 (17.6)	2 (2.7)
Pi < 2.0 (n = 26)	15 (20.3)	5 (6.7)	6 (8.1)

Values are presented as number (%). Pi: perfusion index, SpHb: noninvasive hemoglobin, Lab-Hb: laboratory hemoglobin.

and Lab-Hb ($P = 0.042$).

DISCUSSION

The aim of this study was to determine the agreement of SpHb against Lab-Hb in C-sec causing massive and acute intraoperative blood loss, and our findings showed that SpHb had a good correlation with Lab-Hb.

In practice, a laboratory Hb is usually considered as the gold standard for measuring intraoperative Hb. However, laboratory Hb measurements only provide intermittent and delayed results in the clinical condition of massive and acute blood loss due primarily to its turnaround time. Therefore, frequent laboratory evaluation is required to detect the decrease of Hb level during an operation, which is not only time-consuming but also labor intensive, requiring specific trained personnel.

On the other hands, continuous noninvasive Hb monitoring can obtain real-time change of Hb. The accuracy of noninvasive Hb monitoring has been reported in several surgeries presenting significant intraoperative bleeding [12–14]. In parturient undergoing C-sec, noninvasive Hb was also evaluated at pre- and post-operative periods and it did not seem to be as accurate as the laboratory one [15,16]. Different from the previous studies, this study enrolled the pregnant population with placenta previa; therefore, intraoperative massive and rapid blood loss was expected. In addition, we focused the correlation between the two variables only during the intraoperative period, when substantial blood loss is expected to occur and immediate transfusion can be initiated.

In our results, the correlation coefficient between the two measurements was relatively high; therefore, SpHb may be valuable in continuous monitoring Hb trend. However, a high correlation coefficient does not imply that there is a good agreement between the two methods. Thus, the focus should be more on the results of Bland–Altman analysis. On

average, SpHb was recorded to be 0.3 g/dl higher than Lab-Hb according to the Bland–Altman plot. This small difference did not seem to be a problem in clinical setting, especially in situations where transfusion is appropriate. However, the limits of agreements were not narrow enough to be acceptable for clinical purpose, such as detection of anemia triggering RBC transfusion. In addition, we found a slight negative trend of the differences, proportional to the increment of the measurement. When we regressed the differences of the two Hb measurements on their average, the standardized coefficient for the mean differences was -0.3 . Although the coefficient was very small and only 15 datasets existed in the mean Hb level of < 10 g/dl, it is important to keep in mind that SpHb could be measured higher than the Lab-Hb in anemic conditions. Even though SpHb and Lab-Hb were calibrated initially with each other, there could be a deviation between SpHb and Lab-Hb in anemic conditions after substantial blood loss. If SpHb was overestimated in the anemic population, a decision to initiate RBC transfusion could be delayed. It was reported that adding SpHb to the standard laboratory Hb measurement decreased the volume of transfusion in surgeries with high blood loss, such as several neurosurgeries [17], and SpHb was remarkably consistent with Lab-Hb in trauma patients with Hb under 8 g/dl [18]. However, it remains controversial whether SpHb is enough to detect severe anemia or to suggest the need for RBC transfusion solely [10,19,20].

In our patients, Pi measured simultaneously with SpHb did not change drastically during surgery; although, there were individual differences. Pi should be considered to evaluate the accuracy of SpHb, and it has been reported that SpHb is more accurate in patients with $Pi > 2.0$ [21–23]. In our study, SpHb also tended to be more consistent with Lab-Hb in the subgroup of $Pi \geq 2.0$. Thus, interpretation of SpHb should be made with caution in clinical conditions with decreasing Pi, such as hypothermia, vasoactive drugs, or increased sympathetic tone.

When we use continuous and noninvasive Hb monitoring, it should be confirmed whether it is really helpful in the transfusion practice, such as timely initiation of transfusion, avoidance of over-transfusion, and proper amount of transfusion. In the present study, transfusion was performed based on Lab-Hb; thus, we could not reveal the effectiveness of SpHb on the transfusion practice. Further study is necessary to evaluate this issue.

According to the previous report, Hb of venous blood was higher than that of arterial blood [24]. In addition, Hb measured by co-oximetry method was shown to be higher than Hb measured in central laboratory (reported as complete blood counts) [25]. In our study, Lab-Hb was measured using arterial blood via the co-oximetry method. Depending on which method is set as the gold standard of Hb measurement, there can be larger or smaller differences between SpHb and Lab-Hb. It should be considered carefully that a larger bias is likely to cause concern in clinical diagnosis of anemia triggering RBC transfusion.

This study had several limitations. First, the datasets of SpHb and Lab-Hb in anemic conditions, especially 8 g/dl or less, were definitely lacking. When anemia was noticed and massive bleeding continued, RBC transfusion was initiated instantly to our participants. Therefore, severe anemic condition was rare in this study. Second, we did not consider the effects of vasoconstrictors used during the surgery. After spinal anesthesia, several parturients presented hypotensive events, when it was treated with phenylephrine or ephedrine with fluid administration. In addition, spinal anesthesia was converted to general anesthesia in two patients during the surgery. Thus, dynamic change probably occurred. As a result, the range of observed perfusion index on noninvasive Hb monitoring was from 0.47 to 9.2 in this study.

In conclusion, our study showed that SpHb monitoring had a good correlation with Lab-Hb. A small mean difference between noninvasive SpHb and Lab-Hb might not be clinically significant; however, the limits of agreements was not narrow. In particular, SpHb could be overestimated in the anemic population. Based on our results, further studies investigating the accuracy and precision of noninvasive Hb monitoring should be performed in parturients presenting Hb below 10 g/dl.

ACKNOWLEDGMENTS

This research was supported by the 2017 Research Grant funded by a Korean Society of Obstetric Anesthesiologists.

CONFLICTS OF INTEREST

No potential conflict of interest relevant to this article was reported.

ORCID

Hyunsung Kim: <https://orcid.org/0000-0001-7681-9280>

Sang-Hwan Do: <https://orcid.org/0000-0001-5452-4166>

Jung-Won Hwang: <https://orcid.org/0000-0002-0887-6889>

REFERENCES

- García-Soler P, Camacho Alonso JM, González-Gómez JM, Milano-Manso G. Noninvasive hemoglobin monitoring in critically ill pediatric patients at risk of bleeding. *Med Intensiva* 2017; 41: 209-15.
- Xu T, Yang T, Kim JB, Romig MC, Sapirstein A, Winters BD. Evaluation of noninvasive hemoglobin monitoring in surgical critical care patients. *Crit Care Med* 2016; 44: e344-52.
- Martin JR, Camp CL, Stitz A, Young EY, Abdel MP, Taunton MJ, et al. Noninvasive hemoglobin monitoring: a rapid, reliable, and cost-effective method following total joint replacement. *J Bone Joint Surg Am* 2016; 98: 349-55.
- Galvagno SM Jr, Hu P, Yang S, Gao C, Hanna D, Shackelford S, et al. Accuracy of continuous noninvasive hemoglobin monitoring for the prediction of blood transfusions in trauma patients. *J Clin Monit Comput* 2015; 29: 815-21.
- Phillips MR, Khoury AL, Bortsov AV, Marzinsky A, Short KA, Cairns BA, et al. A noninvasive hemoglobin monitor in the pediatric intensive care unit. *J Surg Res* 2015; 195: 257-62.
- Moore LJ, Wade CE, Vincent L, Podbielski J, Camp E, Junco DD, et al. Evaluation of noninvasive hemoglobin measurements in trauma patients. *Am J Surg* 2013; 206: 1041-7.
- Bhat A, Upadhyay A, Jaiswal V, Chawla D, Singh D, Kumar M, et al. Validity of non-invasive point-of-care hemoglobin estimation in healthy and sick children-a method comparison study. *Eur J Pediatr* 2016; 175: 171-9.
- Bourke M, Cooley SM, McMorro R, McAuliffe FM. Noninvasive hemoglobin measurement in pregnancy. *Int J Gynaecol Obstet* 2014; 125: 78-9.
- Pyregov AV, Ovechkin Alu, Petrov SV. Noninvasive total hemoglobin monitoring based on multiwave spectrophotometry in obstetrics and gynecology. *Anesteziol Reanimatol* 2012; (6): 36-9.
- Huang PH, Shih BF, Tsai YF, Chung PC, Liu FC, Yu HP, et al. Accuracy and trending of continuous noninvasive hemoglobin monitoring in patients undergoing liver transplantation. *Transplant Proc* 2016; 48: 1067-70.
- Suehiro K, Joosten A, Alexander B, Cannesson M. Continuous noninvasive hemoglobin monitoring: ready for prime time? *Curr Opin Crit Care* 2015; 21: 265-70.

12. Berkow L, Rotolo S, Mirski E. Continuous noninvasive hemoglobin monitoring during complex spine surgery. *Anesth Analg* 2011; 113: 1396-402.
13. Park YH, Lee JH, Song HG, Byon HJ, Kim HS, Kim JT. The accuracy of noninvasive hemoglobin monitoring using the radical-7 pulse CO-Oximeter in children undergoing neurosurgery. *Anesth Analg* 2012; 115: 1302-7.
14. Kim SH, Choi JM, Kim HJ, Choi SS, Choi IC. Continuous noninvasive hemoglobin measurement is useful in patients undergoing double-jaw surgery. *J Oral Maxillofac Surg* 2014; 72: 1813-9.
15. Butwick A, Hilton G, Carvalho B. Non-invasive haemoglobin measurement in patients undergoing elective Caesarean section. *Br J Anaesth* 2012; 108: 271-7.
16. Skelton VA, Wijayasinghe N, Sharafudeen S, Sange A, Parry NS, Junghans C. Evaluation of point-of-care haemoglobin measuring devices: a comparison of Radical-7TM pulse co-oximetry, HemoCue[®] and laboratory haemoglobin measurements in obstetric patients. *Anaesthesia* 2013; 68: 40-5.
17. Awada WN, Mohmoued MF, Radwan TM, Hussien GZ, Elkady HW. Continuous and noninvasive hemoglobin monitoring reduces red blood cell transfusion during neurosurgery: a prospective cohort study. *J Clin Monit Comput* 2015; 29: 733-40.
18. Gamal M, Abdelhamid B, Zakaria D, Dayem OAE, Rady A, Fawzy M, et al. Evaluation of noninvasive hemoglobin monitoring in trauma patients with low hemoglobin levels. *Shock* 2018; 49: 150-3.
19. Saito J, Kitayama M, Oishi M, Kudo T, Sawada M, Hashimoto H, et al. The accuracy of non-invasively continuous total hemoglobin measurement by pulse CO-Oximetry undergoing acute normovolemic hemodilution and reinfusion of autologous blood. *J Anesth* 2015; 29: 29-34.
20. Baulig W, Seifert B, Spahn DR, Theusinger OM. Accuracy of non-invasive continuous total hemoglobin measurement by Pulse CO-Oximetry in severe traumatized and surgical bleeding patients. *J Clin Monit Comput* 2017; 31: 177-85.
21. Park SG, Lee OH, Park YH, Shin HY, Kang H, Baek CW, et al. The changes of non-invasive hemoglobin and perfusion index of Pulse CO-Oximetry during induction of general anesthesia. *Korean J Anesthesiol* 2015; 68: 352-7.
22. Miller RD, Ward TA, McCulloch CE, Cohen NH. Does a digital regional nerve block improve the accuracy of noninvasive hemoglobin monitoring? *J Anesth* 2012; 26: 845-50.
23. Nguyen BV, Vincent JL, Nowak E, Coat M, Paleiron N, Gouny P, et al. The accuracy of noninvasive hemoglobin measurement by multiwavelength pulse oximetry after cardiac surgery. *Anesth Analg* 2011; 113: 1052-7.
24. Yang ZW, Yang SH, Chen L, Qu J, Zhu J, Tang Z. Comparison of blood counts in venous, fingertip and arterial blood and their measurement variation. *Clin Lab Haematol* 2001; 23: 155-9.
25. Carabini LM, Navarre WJ, Ault ML, Bebawy JF, Gupta DK. A comparison of hemoglobin measured by co-oximetry and central laboratory during major spine fusion surgery. *Anesth Analg* 2015; 120: 60-5.